

**HYNAUT**

海氏海诺®

Qingdao Hainuo Biological Engineering Co., Ltd.

青岛海诺生物工程有限公司

Qingdao Hainuo Biological Engineering Co., Ltd. is a biotechnology enterprise integrating R & D, production and sales. It is affiliated to Qingdao Haishi Hainuo Group. The Group was founded in 2004 and now has more than 20 subsidiary corporations. Our headquarter is located in Jiangshan Industrial Park in Laixi Qingdao, with superior geographical location and convenient transportation, covering an area of 330,000 square meters, which is adjacent to Jiangshan Lake, Qingrong Light Rail, major highways, Qingdao International Airport and Qingdao Port, and It's one step away from Huashan International Golf Course.



# HYNAUT

海氏海诺®

We have several modern workshops and logistics storage facilities that meet international standards, equipped with advanced production facilities and high-precision testing instruments. There are nearly 2,000 employees including experts, professors and licensed pharmacists from the Chinese Pharmaceutical Association. Technical support and collaboration are provided by many scientific research institutions and companies from USA, Germany and Taiwan, China.



# Clean Workshop



**HYNAUT**  
海氏海诺®



**Production  
Line**

## Detecting instrument



## Experts of the Thousand Talents Plan



Professor Hui Miyu's patented products, B-HA (active hyaluronic acid) toothpaste and nasopharyngeal spray, broad-spectrum anti-bacterial anti-inflammatory, have been patented in the United States.

Professor Ning Xin's main research areas are nonwoven materials and equipment, hygiene and medical supplies, industrial textile composites, polymer drug release systems, environmental engineering functional fibers and materials.



Professor Jiang Guohui has presided over the topic of using nuclear technology to study new special effects of herbal active ingredients, which has received the support of the National Natural Science Foundation of China and the National Science Research Project Fund, and he has rich experience in pharmacology, toxicology, molecular biology and research and development of new drugs.



## Chen Pu academician



Academician of the Canadian Academy of Engineering and National Chief Professor of Nanobiomaterials in Canada. The small nucleic acid anti-tumor drug research and development project led by Academician Chen is at the international advanced level in the field of nanobiomedicine technology. We cooperated with Academician Chen to establish Qingdao TC-Norder Biotechnology Co., Ltd. The company's project is progressing smoothly. It is planned to set up a workstation in Jiangshan and will achieve 40 million sales revenue in 2020.



Seeram Ramakrishna academician

---



Foreign academician of the Royal Academy of Engineering, Indian National Academy of Engineering, Academician of the Singapore Academy of Engineering. His research direction is the processing and surface functionalization of polymer composite materials, biocomposite materials, and electrospinning nanofibers. He has conducted original research in many scientific fields and made great contributions in the field of nanotechnology.

## Guan huashi academician

---

Academician of the Chinese Academy of Engineering, an expert in China's marine medicine, an advocate of China's "blue drug bank", and a major developer of treating Alzheimer's disease with marine organisms. We cooperated with Institute of Marine Medicine and Biology of Qingdao Ocean University to set up Haisheng Health Technology Co., Ltd to jointly develop marine medicine products.



Haishi Hainuo Group has been repeatedly awarded as one of the "Top Three Enterprises in China's Band-Aid Industry", and cartoon Band-Aids pioneered the domestic band-aid industry. The company's main products are: face masks, wound stickers, medical infusion stickers, pressure-sensitive adhesive tape, medical disinfection cotton balls, medical disinfection cotton pads, trauma emergency kits, syringes, infusion sets, early pregnancy test paper, ovulation test paper, foot stickers, warm patch, medical antipyretic stickers, gauze bandages, disinfectant, dental floss sticks, hyaluronic acid toothpaste, medical care pads, ice pads, condoms, etc.



# HYNAUT

海氏海诺®

The group has successively won the honorary titles of "Contract-keeping and Credit-honoring Enterprise of Qingdao", "Excellent Private Enterprise of Qingdao", "Excellent Award of Shandong Famous Enterprise Expo", "Outstanding Brand in Medical Dressing Industry of China", "Excellent Private Enterprise of China".

Chairman Mr. Ma Zhaohui won the "Trailblazing Entrepreneur Award of Qingdao".

In 2010, the "HAINUO" trademark was rated as "Famous Trademark of Shandong Province", and the HAINUO Trauma Emergency Kit was rated as "Famous product of Shandong Province".

In 2012, the company was rated as "Laixi Outstanding Welfare Enterprise", "Top Three Most Innovative and Competitive Enterprises in China's Medical Device Industry" and "Qingdao High-tech Enterprise".

In 2013, the company was rated as "National Welfare Enterprise Demonstration Unit" by the China Social Welfare Association and "Qingdao Medical Dressing Engineering Technology Research Center". "Hainuo Hainuo" trademark was rated as "Shandong Famous Trademark", 2015 "HAINUO" brand was rated as "China Famous Brand".

In 2016, the company was named "Invisible Champion" Enterprise of Qingdao Industrial Enterprise.

In 2017, the company's product "Sterile patches" was rated as "Qingdao Specialized New Product", and the company was rated as Qingdao Specialized Special New Model Enterprise.



In 2008, the company reached a cooperation with American Warner Company, making the world-famous cat and mouse cartoon image become the endorsement of Hainuo products. In 2013, on the basis of "TOM and JERRY", the company successfully signed the "Superman" image to become the company's product image endorsement. Products of Hainuo have entered into thousands of households.



Qingdao Hainuo Biological Engineering Co., Ltd is on the

Name List of Medical Devices and Supplies Companies with Certification/Authorization from other Countries



**CHINA CHAMBER OF COMMERCE FOR  
IMPORT & EXPORT OF MEDICINES & HEALTH PRODUCTS**

<http://www.cccmhpie.org.cn/>



**中国医药保健品进出口商会**  
服务 产业链 | 助力 国际化

English 登陆 | 注册

请输入关键词进行搜索



开具不可抗力相关事实性证明

取得国外认证和注册企业查询

首页

关于商会

新闻中心

行业服务

权威发布

商会会刊

企业风采

会员之家

加入商会

取得国外标准认证或注册的医疗物资和非医用口罩生产企业检索

青岛海诺

检索

企业名称 (中文)

企业名称 (英文)

产品类别

统一社会信用代码

国外注册认证情况

青岛海诺生物工程有限公司

Qingdao Hainuo Biological Engineering Co., Ltd

医用口罩

913702857180717488

欧盟CE

Medical Face Mask

EU CE

Qingdao Hainuo Biological Engineering Co., Ltd. under the group is a national white-listed enterprise for anti-epidemic materials, a special credit subsidy unit for financial credit granted by banks, and a white-listed enterprise for face mask approval by the State and Provincial Drug Administration.

## 中华人民共和国商务部 中华人民共和国海关总署 国家市场监督管理总局

### 公告

2020年第12号

#### 关于进一步加强防疫物资出口质量监管的公告

在全球疫情持续蔓延的特殊时期，为更有效支持国际社会共同应对全球公共卫生危机，现就进一步加强防疫物资质量监管、规范出口秩序有关措施公告如下：

一、加强非医用口罩出口质量监管。自4月26日起，出口的非医用口罩应当符合中国质量标准或国外质量标准。

商务部确认取得国外标准认证或注册的非医用口罩生产企业清单（中国医药保健品进出口商会网站 [www.cccmhpie.org.cn](http://www.cccmhpie.org.cn) 动态更新），市场监管总局提供国内市场查处的非医用口罩质量不合格产品和企业清单（市场监管总局网站 [www.samr.gov.cn](http://www.samr.gov.cn) 动态更新），非医用口罩出口企业报关时须提交电子或书面的出口方和进口方共同声明（参考附件1），确认产品符合中国质量标准或国外质量标准，进口方接受所购产品质量标准且不予于医

#### 权威发布：口罩等防疫物资出口企业“黑名单”出炉！附2235家“白名单”企业清单！

2020-04-26 17:08:00 国际贸易 口罩

昨天（4月25日）傍晚，商务部、海关总署和国家市场监督管理总局紧急发布 2020年第12号公告，宣布进一步升级对包括非医用口罩在内防疫物资出口质量监管措施！

消息一出，业界哗然，这宣告了以瞒报，漏报，夹带和冲关等不合规口罩出口模式，已经完全行不通了！合法合规和产品合格将是口罩出口唯一出路！

在这份最新的12号公告中，三部委特别给出了一份震惊业界的白名单和黑名单！这基本可以看作是合格口罩出口企业的一本通行证，也是伪劣口罩生产企业的一份通缉令！

263	辽械注准20202140108	一次性使用医用口罩	辽宁金凤凰服饰有限公司
264	辽械注准20202140118	一次性使用医用口罩	浩宁实业（沈阳）有限公司
265	鲁械注准20152640334	普通医用口罩	山东爱达医用制品有限公司
266	鲁械注准20152640428	一次性医用口罩	临沂康利医疗器械有限公司
267	鲁械注准20162640455	一次性使用医用口罩	德州康迪医疗用品有限公司
268	鲁械注准20162640461	医用一次性口罩	山东昂扬医疗科技有限公司
269	鲁械注准20162640494	一次性使用医用口罩（非外科用	青岛海诺生物工程有限公司
270	鲁械注准20162640492	医用口罩	山东华晨医疗器械有限公司
271	鲁械注准20172640395	一次性使用医用口罩	青岛科美生物工程有限公司
272	鲁械注准20172640652	一次性使用医用口罩（非外科用	山东朱氏药业集团有限公司
273	鲁械注准20172640889	一次性使用医用口罩	山东省聚成医疗器械有限公司
274	鲁械注准20182140392	一次性使用医用口罩	山东创新医疗器械科技有限公司
170	鲁械注准20142140149	医用外科口罩	威海鸿宇无纺布制品有限公司
171	鲁械注准20152640383	医用外科口罩	青岛盛久医疗用品有限公司
172	鲁械注准20162640237	医用外科口罩	山东创新医疗器械科技有限公司
173	鲁械注准20162640388	医用外科口罩	德州康迪医疗用品有限公司
174	鲁械注准20162640493	一次性使用医用外科口罩	青岛海诺生物工程有限公司
175	鲁械注准20162640602	医用外科口罩	青岛上康利医疗器械有限公司
176	鲁械注准20172640005	医用外科口罩	日照三奇医疗卫生用品有限公司
177	鲁械注准20172640350	医用外科口罩	山东省聚成医疗器械有限公司
178	鲁械注准20172640500	一次性使用医用外科口罩	山东九尔实业集团有限公司



We have achieved long-term strategic cooperation with more than 200,000 pharmacies and a large number of chain convenience stores in China. Our products not only enjoy high popularity and reputation in domestic market, but also exported to overseas, favored by European, American and other customers all over the world. In recent years, as the Group grows in size, our products, production lines, research& development teams are all growing. We are always at the forefront of the industry, committed to providing the world's people with the most cutting-edge products and best service.




## Partners: China's top 100 chain pharmacies



The company has undertaken many international marathons and equestrian competitions, and strives to promote the integration and development of sports, culture and technology.





After the outbreak of the coronavirus COVID-19, the company invested more than 80 million yuan and added a number of new production lines for disposable medical face masks, KN95 protective masks, alcohol & iodine wipes and PE gloves production lines. Currently, the daily output of face masks reaches 3.6 million.

## Related Reports from CCTV, Xinhua Net, Qingdao News





The Group has donated to Wuhan & Qingdao Medical Team, Pan Ki-moon Foundation and other domestic and foreign anti-epidemic first-line donations including protective supplies, alcohol disinfection products and daily necessities, with a total value of more than 10 million yuan.



# Face mask Series

- HYNAUT brand.
- Direct sales from the factory to ensure high quality.
- Direct contract signing to ensure safe transactions.
- CE / FDA / TGA and other qualifications are complete.

Various product test reports are available.



## Disposable Face Mask

non-woven + melt-blown + non-woven

- *Personal Protective*
- *Medical EN14683 type I*
- *Medical EN14683 type IIR*



Packing	Carton L*W*H	Gross Weight	Net Weight	Volume
1pcs/bag*200bags=200pcs/ctn	73.5*47*31	8.5	7.2	0.107
50pcs/box*40boxes=2000pcs/ctn	52*39*42.5	9.05	8	0.086



## KN95 Protective Face Mask ( Willow Leaf )

- 4-ply design
- non-woven + melt-blown x2 + non-woven
- Four layers of protection, multiple filtration, filtration efficiency  $\geq 95\%$ .  
Willow-leaf cut, fits the face more comfortably.



Packing	Carton L*W*H	Gross Weight	Net Weight	Volume
1pcs/bag*200bags=200pcs/ctn	29.5*25.5*22.5	1.9	1.55	0.017





## KN95 Protective Face Mask ( Folded )

- 4-ply design
- non-woven + melt-blown x2 + non-woven
- Four layers of protection, multiple filtration, filtration efficiency  $\geq 95\%$ .  
Three-dimensional design, very comfortable to wear.



Packing	Carton L*W*H	Gross Weight	Net Weight	Volume
1pcs/bag*200bags=200pcs/ctn	34*20.5*22.5	1.9	1.55	0.016



## KN95 Protective Face Mask ( For kids )

- 4-ply design
- non-woven + melt-blown x2 + non-woven
- Four layers of protection, multiple filtration, filtration efficiency ≥95%.



Packing	Carton L*W*H	Gross Weight	Net Weight	Volume
1pcs/bag*200bags=200pcs/ctn	31.5*19.0*21.5	1.6	1.35	0.013

# Qualification



**营业执照**

(副本) 1-1

统一社会信用代码  
913702857180717488

扫描二维码  
了解更多企业信息  
请登录国家企业信用信息公示系统  
网址: <http://www.gsxt.gov.cn>

名称	青岛海诺生物工程有限公司	注册资本	贰仟万元整
类型	有限责任公司(自然人投资或控股)	成立日期	2000年05月29日
法定代表人	麻兆晖	营业期限	2000年05月29日至 年 月 日
经营范围	普通货物运输, 专用货物运输(集装箱)(以上项目不含危险品及违禁品, 并依据道路运输管理局颁发的许可证从事经营活动), I类、II类、III类医疗器械的生产、销售; 化妆品、消毒用品的生产、销售(危险品除外); 劳保用品的生产、销售; 日用百货销售; 口罩、清洁用品、驱蚊用品、口腔用品、晕车贴、皮肤清洁用品、热敷贴、药盒、呼吸贴的生产、销售; 生物技术研究、开发; 以上货物、技术进出口。(依法须经批准的项目, 经相关部门批准后方可开展经营活动)		
登记机关	莱西市行政审批服务局		
	2020年03月23日		

市场主体应当于每年1月1日至6月30日通过国家企业信用信息公示系统报送公示年度报告

国家企业信用信息公示系统网址:

国家市场监督管理总局监制


**医疗器械生产许可证**

许可证编号鲁食药监械生产许 20120047 号

企业名称	青岛海诺生物工程有限公司	生产地址	1. 青岛市莱西市姜山镇工业园广东路1号; 2. 青岛市莱西市姜山镇工业园海氏海诺新工业园区1号
法定代表人	麻兆晖	生产范围	II类: 6840 体外诊断试剂, 6864 医用卫生材料及敷料※※
企业负责人	刘宝玉	发证部门	山东省食品药品监督管理局
住所	青岛市莱西市姜山镇工业园	有效期限	至 2021 年 11 月 15 日
		发证日期	2018 年 05 月 08 日

国家食品药品监督管理总局制

国家食品药品监督管理总局制



## 对外贸易经营者备案登记表

备案登记表编号：

进出口企业代码：

经营者中文名称	青岛海诺生物工程技术有限公司		
经营者英文名称	QINGDAO HAINUO BIOLOGICAL ENGINEERING CO.,LTD		
组织机构代码	718071748	经营者类型 (由备案登记机关填写)	有限责任公司
住 所	青岛市莱西市莱山镇工业园		
经营场所 (中文)	青岛市莱西市莱山镇工业园		
经营场所 (英文)	Jiangshan International Industrial Area of Laixi Qingdao		
联系电话	053286463333	联系传真	053286460000
邮政编码	266603	电子邮箱	hn@hainuocn.com
工商登记注册日期	2000-5-20	工商登记注册号	370102100000017

依法办理工商登记的企业还须填写以下内容

企业法定代表人姓名	麻兆晖	有效证件号	
注册资金	贰仟万元	(折美元)	

依法办理工商登记的外国（地区）企业或个体工商户（独资经营者）还须填写以下内容

企业法定代表人 个体工商户负责人姓名	有效证件号
企业资产/个人财产	(折美元)

备注
----

填表前请认真阅读背面的条款，并由企业法定代表人或个体工商户负责人签字。



2013 年 11 月 06 日



### CONFIRMATION OF FDA REGISTRATION

Registration No.300880550

Dear Official Correspondent:  
This document provides notification of the registration number assigned to your establishment.

Establishment: QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD.  
Address: JIANGSHAN INDUSTRIAL AREA OF JIANGSHAN TOWN, LAIXI, QINGDAO, SHANDONG, 266603, CHINA  
Listing Number: See Appendix

*Helen Han*

General Manager  
Email: ce-086@hotmail.com  
Web: www.cmc.com



Validity: Dec.31,2020

**Conclusion:**  
The applicant makes an initial representation or declaration, not done in order for representation and acceptance to any person or entity other than the named certifier below. Check entries on liability to any person or entity in connection with this company. The list, kind and class information does not change quantities of registration, nor does it list, kind and class information. Information is considered as complete. Check entries on liability to any person or entity in connection with this company. Please send back: FDA registration certificate for manufacturer's information for future and correct products with FDA, which does not mean that products have correct quality specifications unless the manufacturer certifies on the quality certificate.

FDA Official Website: <https://www.accessdata.fda.gov/cdrh/cdrh/rdm/cdrh.cfm>



### CONFIRMATION OF FDA REGISTRATION

Registration No.300880550

Appendix:

Listing Number	Product Submitter Number/Type	Product Code(s)	Device Name(s)	Proprietary Name
D03394	Reference: Disinfectant	LAK	Put, alcohol, device disinfectant	ALCOHOL WIPES, ALCOHOL PREP PAD
D03044	Example	K30	APPLICATOR, ABSORBENT TIPPED, NONSTERILE	COTTON SWAB
D10084	Example	NAC	Cleaning, wound, hydrophilic	WOUND DRESSING/ADHERIVE, GAUZE PAD, UNDERPAD, COTTON BALL
D10300	Example	NAB	Cover (glove, glove/sterile) the manual use	RUDDIE SWAB
D10142	Example	LBT	Disinfectant, medical device	
D14556	Example	KGX	Tape and bandage, adhesive	ADHERIVE BANDAGE, WOUND PLASTER
D14143	Example	BS	FLORA, ORINAL	TOILET PAPER
D10970	Reference: Disinfectant	QKA	Face mask (surgical) required for general public/healthcare personnel per IEC guidance	DISPOSABLE PROTECTIVE FACE MASK
D10395	Reference: Disinfectant	ORR	Mask, surgical, non-sterile	FACE MASK
D10194	Reference: Disinfectant	ORR	Face and lip-wobster dress	FIRST AID KIT
D14667	Example	FCR	BANDAGE, ELASTIC	ELASTIC BANDAGE/GAUZE BANDAGE
D14668	Example	BD	PACK, HOT OR COLD, DISPOSABLE	WARM PATCH, COLD PATCH, FEVER RELIEF PATCH, COOL GEL PATCH/ GEL
D09170	Example	NXA	MASK, REVENGING	

**Conclusion:**  
The applicant makes an initial representation or declaration, not done in order for representation and acceptance to any person or entity other than the named certifier below. Check entries on liability to any person or entity in connection with this company. The list, kind and class information does not change quantities of registration, nor does it list, kind and class information. Information is considered as complete. Check entries on liability to any person or entity in connection with this company. Please send back: FDA registration certificate for manufacturer's information for future and correct products with FDA, which does not mean that products have correct quality specifications unless the manufacturer certifies on the quality certificate.

FDA Official Website: <https://www.accessdata.fda.gov/cdrh/cdrh/rdm/cdrh.cfm>

### EU Declaration of Conformity

Manufacturer: QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD  
NO.1 GUANGDONG ROAD,JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA  
Tel: +86-532-8946333/8946444

SRN: /

European Representative: CMC Medical Devices & Drugs S.L.  
Paseo Lengo N° 18, CP 29008, Málaga, Spain

Product Name: Disposable Medical Face Mask  
Product Type: EN14683:2019 TYPE I  
Product Code: C207  
Specification: 17.5cm\*5cm  
GMDN Code: 12-458

Classification (MDD, Annex IX), Class I, Rule 1  
Conformity Assessment Route: Annex IX of MDD 93/42/EEC

We hereby declare that the above mentioned products meet the transposition into national law, the provisions of the following EU Regulation and Standards. All supporting documentations are retained under the provision of the manufacturer.

The manufacturer is exclusively responsible for the declaration of conformity.

General applicable regulations: directives  
Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC of June 1993 concerning medical devices (MDD 93/42/EEC), Amended by DIRECTIVE 2007/47/EC of 5 September 2007

Applied standards, common specification, guidance  
EN 14683:2019+AC:2019, EN ISO 15223-1:2016, EN 1041:2008+A1:2013, EN ISO 14879:2012, EN 62368-1:2018+AC:2019, ISO 10993-1:2018, ISO 10993-5:2009, ISO 10993-10:2010, MDCC 2019-15

Signature: *Helen Han*  
Name: Helen Han  
Position: General Manager  
Place/Date: Qingdao 6/9/2020  
File No.: HNC01-01, ver A0

### EC REP CERTIFICATE

CMC MEDICAL DEVICES & DRUGS SL  
NO. CMC/CE/2020/03042020.5

CONFIRMED THAT CMC MEDICAL DEVICES & DRUGS S.L. IS THE European Authorized Representative of

QINGDAO HAINUO BIOLOGICAL ENGINEERING CO LTD  
NO.1 GUANGDONG ROAD, JIANGSHAN INDUSTRIAL ZONE LAIXI CITY QINGDAO

The certificate remains valid until the expiration agreement of EC REP, manufacturing conditions, the quality system or relevant legislation are changed. The validity is conditioned by positive results of periodic surveillance audits.

The product liability rests with the manufacturer in accordance with applicable directive and standard, after fulfilling of the relevant EU registration requirements, the manufacturer shall affix relevant CE marking to all already mentioned models of the medical device.

Complies with the applicable essential requirements of the council directive 93/42/EEC on medical devices as amended.

The products in Annex 1 was registered in Spanish MOH with number RPS/433/2020

Issued on: 03/04/2020

Valid until: 02/04/2021

Authorized Representative  
CMC Medical Devices & Drugs SL

[www.cmcmedicaldevices.com](http://www.cmcmedicaldevices.com)

### EC REP CERTIFICATE

ANNEX 1 Medical Device Products  
Disposable Medical Face Mask

QINGDAO HAINUO BIOLOGICAL ENGINEERING CO LTD

[www.cmcmedicaldevices.com](http://www.cmcmedicaldevices.com)

USA FDA

EU CE



**Public Summary**  
**Summary for ARTG Entry:** 334843 Australia and New Zealand Health Products Limited - Mask, surgical, single use

**ARTG entry for:** Medical Device Included Class 1  
**Sponsor:** Australia and New Zealand Health Products Limited  
**Postal Address:** Suite 802 Level 8 167 - 169 Queen Street, Melbourne, VIC, 3000 Australia  
**ARTG Start Date:** 22/04/2020  
**Product category:** Medical Device Class 1  
**Status:** Active  
**Approval area:** Medical Devices

**Conditions**

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on its ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

**Manufacturers**  
**Name:** Qingdao Hainuo Biological Engineering Co Ltd  
**Address:** Jiashan Industrial Area of Jinshan Town Laixi Qingdao, Shandong, 266600, China

**Products**

**1. Mask, surgical, single use**

**Product Type:** Single Device Product **Effective date:** 22/04/2020

**GMON:** 35177 Mask, surgical, single use  
**Intended purpose:** This mask is to assist in the reduction of the spreading of germs and bacteria.  
**Specific Conditions:**

© Commonwealth of Australia. This work is copyright. You are not permitted to re-transmit, distribute or commercialise the material without obtaining prior written approval from the Commonwealth. Further details can be found at <http://www.tga.gov.au/about/webste-copyright.htm>.

### 시험성적서

성적서번호 : MY18-00275

○ 시험결과

시험검사항목	단 위	기 준	시험-검사 결과	항목판정	
성상	-	육안관찰	이상없음	적합	
형상	mm	가로	마스크의 장변을 가로로 정의함	130	적합
		세로	가로에 수직방향을 세로로 정의함	157	
		깊이	머리끈의 장변을 깊이로 정의함	최 : 167 무 : 158	
머리끈 폭	깊이의 수직방향을 폭으로 정의함	5.2			
색도	-	색소	관찰하여 색을 나타내지 않음	색 나타내지 않음	적합
		산 또는 알칼리	홍색을 나타내지 않음(메틸프탈레인 시약) 적색을 나타내지 않음(메틸오렌지 시약)	홍색, 적색 나타내지 않음	적합
		형광	자외선(350~370 nm)에서 형광을 나타내지 않음	형광 없음	적합
		포름알데히드	전기성 형광증폭제 시험에서 형광 유 무	진하지 않음	적합
인장강도	N	절단하중(N) 평균(3회)이 10 N 이상	평균값	39.2	적합
			본품	40.9	
인면부흡기저항	Pa	6개 각각의 결과가 70 Pa (KF94) 이하	본품	42.0	적합
			전처리	41.4	
			본품	51.1	
포집효율(NaCl)	%	본품 3개, 전처리 3개 결과가 94 % (KF94) 이상	본품	99.1	적합
			전처리	99.2	
			본품	98.8	
포집효율(파라핀오일)	%	본품 3개, 전처리 3개 결과가 94 % (KF94) 이상	본품	99.2	적합
			전처리	98.2	
			본품	98.1	
			전처리	98.3	

The National Personal Protective Technology Laboratory (NPPTL)

NIOSH > NPPTL > Respirator Assessments to Support COVID-19 Response



NPPTL

What's New on the NPPTL Website +

A to Z Index

For Respirator Users +

Promoting productive workplaces through safety and health research **NIOSH**

## NPPTL Respirator Assessments to Support the COVID-19 Response

Updated May 8, 2020

### International Assessment Results – Not NIOSH-approved

NPPTL has completed International Assessments for the products listed below.

NPPTL makes no representation as to the authenticity of the samples received and assessed. As part of its standard respirator approval process for NIOSH-approved respirators, NPPTL conducts a comprehensive quality assurance review of the quality process and manufacturing site. None of these reviews were conducted during this limited assessment. Further, no certificates of approval were provided with the samples. Therefore, validation of the claims that the product meets a particular international standard cannot be made.

For each model listed, ten respirators were submitted for evaluation. The samples were tested using a modified version of NIOSH Standard Test Procedure (STP) TEB-APR-STP-0059. Only particulate filter efficiency was assessed. The results of these tests are for the sample tested and may or may not be representative of a larger lot or population of similar respirators. The results of any filter penetration test can only be used to provide a check of the product's filter efficiency. No conclusions can be made regarding equivalency to N95 products that are NIOSH approved.

No certificates of approval were provided with the samples received. Therefore, the authenticity of the claims, that the product meets a particular international standard, cannot be validated.

**These assessments are not a part of the NIOSH respirator approval process and will in no way lead to or preclude NIOSH approval through the official approval process.**

Manufacturer	Model Number/Product Line	International Standard	Filtration Efficiency (%)	Test Report	
			Maximum	Minimum	
Purvigor	KN95 Antibacterial Mask	GB2626	69.80	<a href="#">2020-33.1</a>	
<b>Qingdao Hainuo Bioengineering Co., Ltd.</b>	<b>Hynaut KN95 Protective Face Mask</b>	<b>GB2626</b>	<b>98.37</b>	<b>97.98</b>	<a href="#">2020-96.1</a>

**CDC official website announced that our KN95 protective mask passed NPPTL's test**

These assessments were developed as an assessment of the filter efficiency for those respirators represented as certified by an international certification authority, other than NIOSH, to support the availability of respiratory protection to US healthcare workers and other workers due to the respirator shortage associated with COVID-19.

Most of these products have an ear loop design. NIOSH-approved N95s typically have head bands. Furthermore, limited assessment of ear loop designs, indicate difficulty achieving a proper fit. While filter efficiency shows how well the filter media performs, users must ensure a proper fit is achieved.

These results are not to be used by manufacturers, distributors, suppliers, and importers to make claims about their products and/or to influence purchasers.

Manufacturer	Model Number/Product Line	International Standard	Filtration Efficiency (%)	Test Report	
			Maximum	Minimum	
Anhui Baishidun Protective Equipment Co., Ltd.	Baishidun FFP2	EN149, GB2626	53.00	46.10	<a href="#">2020-51.1</a>
Anhui Changli Environmental Protection Technology Co., Ltd.	KN95 Protective Mask	EN149	95.74	94.73	<a href="#">2020-50.1</a>
Purvigor	KN95 Antibacterial Mask	GB2626	74.20	69.80	<a href="#">2020-33.1</a>
<b>Qingdao Hainuo Bioengineering Co., Ltd.</b>	<b>Hynaut KN95 Protective Face Mask</b>	<b>GB2626</b>	<b>98.37</b>	<b>97.98</b>	<a href="#">2020-96.1</a>
Qingdao Maysheng Medical Devices Co., Ltd.	KN95 Protective Mask (Self-Suction Filter Respirator FFP2)	GB2626	99.59	99.30	<a href="#">2020-45.1</a>
Rizhao Sanqi Medical & Health Articles Co., Ltd.	3Q 9505	GB2626	99.39	96.95	<a href="#">2020-29.1</a>
San Jiao	J595-01	EN149	89.00	82.40	<a href="#">2020-104.1</a>

## NPPTL COVID-19 Response: International Respirator Assessment

Manufacturer: Qingdao Hainuo Bioengineering Co., Ltd.  
 Model Tested: Hynaut KN95 Protective Face Mask  
 Date Tested: April 30, 2020

These findings pertain to the respirator Qingdao Hainuo Bioengineering Co., Ltd., Hynaut KN95 Protective Face Mask. The packaging for this product indicates that it meets GB2626-2006 (the Chinese standard for Respiratory Protective Equipment – Non-Powered Air-Purifying Particle Respirator).

Ten respirators were submitted for evaluation. The samples were tested using a modified version of NIOSH Standard Test Procedure (STP) TEB-APR-STP-0059. This modified assessment plan can be found [here](#).

No certificate of approval was provided with the samples received; therefore, the authenticity of the claims cannot be validated.

The maximum and minimum filter efficiency observed was 98.37% and 97.98%, respectively. All ten respirators measured more than 95%.

While the above-listed product classification has similar performance requirements to NIOSH-approved devices, NIOSH does not have knowledge about the sustained manufacturer quality system and product quality control for these products. NIOSH also does not have knowledge about the product's handling and exposures after leaving its manufacturer's control.

In addition, this product is an ear loop design. Currently, there are no NIOSH-approved products with ear loops; NIOSH-approved N95s have head bands. Furthermore, limited assessment of ear loop designs indicate difficulty achieving a proper fit. While filter efficiency shows how well the filter media performs, users must ensure a proper fit is achieved.

**This assessment is not a part of the NIOSH respirator approval process and will in no way lead to or preclude NIOSH approval through the official approval process.** This assessment was developed as an assessment of the filter efficiency for those respirator's represented as certified by an international certification authority, other than NIOSH, to support the availability of respiratory protection to US healthcare workers due to the respirator shortage associated with COVID-19. Only particulate filter efficiency was assessed.

The results provided in this letter are specific to the subset of samples that were provided to NPPTL for evaluation.

These results will be used to update the CDC guidance for [Crisis Capacity Strategies \(during known shortages\)](#).

## NPPTL COVID-19 Response: International Respirator Assessment

### Evaluation of International Respirators

**NPPTL**  
 National Personal Protective  
 Technology Laboratory

**Test:** Modified TEB-APR-STP-0059

**Date Tested:** April 30, 2020

**Report Prepared:** May 3, 2020

**Manufacturer:** Qingdao Hainuo Bioengineering Co., Ltd.

**Item Tested:** Hynaut KN95 Protective Face Mask

**Country of Certification:** China (GB2626-2006)

Pictures have been added to the end of this report.

Filter	Flow Rate (LPM)	Initial Filter Resistance (mmH <sub>2</sub> O)	Initial Percent Leakage (%)	Maximum Percent Leakage (%)	Filter Efficiency (%)
1	85	12.0	1.64	1.64	98.36
2	85	12.7	1.99	1.99	98.01
3	85	12.4	1.63	1.63	98.37
4	85	12.7	1.72	1.72	98.28
5	85	12.5	1.78	1.78	98.22
6	85	12.9	2.02	2.02	97.98
7	85	12.7	1.66	1.66	98.34
8	85	12.7	1.87	1.87	98.13
9	85	13.5	1.68	1.68	98.32
10	85	12.5	1.87	1.87	98.13
<b>Minimum Filter Efficiency: 97.98</b>			<b>Maximum Filter Efficiency: 98.37</b>		

- The test method utilized in this assessment is not the NIOSH standard test procedure that is used for certification of respirators. Respirators assessed to this modified test plan do not meet the requirements of STP-0059, and therefore cannot be considered equivalent to N95 respirators that were tested to STP-0059.
- Respirators tested may not be representative of all respirators with the same certification mark. NIOSH has no control over suppliers and distributors of respirators certified by other national or international parties.
- This assessment is not a confirmation that it conforms with any or all of its specifications in accordance with its certification mark.
- This assessment was not a part of the NIOSH approval program. These results do not imply nor preclude a future approval through the NIOSH respirator approval program.

NPPTL COVID-19 Response: International Respirator Assessment




NPPTL COVID-19 Response: International Respirator Assessment



NPPTL COVID-19 Response: International Respirator Assessment



NPPTL COVID-19 Response: International Respirator Assessment







## 检验报告

TEST REPORT



报告编号: WT201025604

第 1 页, 共 6 页

委托单位: 青岛海诺生物工程有限公司 QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD.  
 委托单位地址: 青岛市莱西市姜山镇工业园广东路1号 NO.1 GUANGDONG ROAD, JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA.  
 样品名称: 防护口罩 KN95 Protective Face Mask  
 型号/规格/等级: C004  
 检验类别: 送样检验  
 检验地点: 龙华实验基地 Longhua Experimental Base

深圳市计量质量检测研究院 (检验检测专用章)  
 批准人: 何行月  
 签发日期: 2020年04月27日  
 签名: 何行月

### 重要声明

Important statement

- 本院是深圳市人民政府依法设置的产品质量监督检验机构, 系社会公益型非营利性技术机构, 为各级政府执法部门进行监督管理提供技术支持和接受社会各界的委托检验。  
 SMQ is a legal non-profit technical institute established by Shenzhen Municipal Government to undertake the quality supervision and inspection of products, and to provide technical support to relevant supervision and administration and also conduct commission test from the society.
- 本院保证检验的科学性、公正性和准确性, 对检验的数据负责, 并对委托单位所提供的样品和技术资料保密。  
 SMQ is committed to assuring the scientificity, impartiality and accuracy of all tests carried out, responsibility for test data gained, and keeping confidential of all test samples and technical documents provided.
- 抽样按照本院程序文件 CX11-01 (抽样程序) 和相应产品的检验细则的规定执行。  
 The sampling should be carried out according to the "sampling procedure" defined in the Procedure Document CX11-01 and relevant testing specifications.
- 报告无主检、审核、批准人签字, 或涂改, 或未盖本院“检验检测专用章”及骑缝章无效, 未经本院许可, 不得部分复印、抽用或篡改本证书/报告内容。  
 Any report having not been signed by relevant responsible engineer, reviewer or authorized approver, or having been altered without authorization, or having not been stamped by both the "Dedicated Testing/Inspection Stamp" and the sealing stamp is deemed to be invalid. Copying or excerpting portion of, or altering the content of the report/certificate is not permitted without the written authorization of SMQ.
- 送样委托检验结果仅对来样有效; 委托检验的样品信息及委托方信息均由委托方填写, 本院不对其真实性及准确性负责。  
 The test results presented in the report apply only to the tested sample. The product information and the applicant information are provided by the customer and SMQ assumes no responsibility for their validity and accuracy.
- 未经检验机构同意, 样品委托人不得擅自使用检验结果进行不当宣传。  
 Any use of SMQ test result for advertisement of the tested material or product must be approved in writing by SMQ.
- 无 CMA 标志的报告, 仅供使用方内部参考, 不具有对社会的证明作用。含粤字编号的 CMA 标志仅适用于产品标准和判定标准。  
 The non-CMA report issued by SMQ is only permitted to be used by the client as internal reference use and shall not be used for public demonstration purpose. CAL logo with symbol "Yue" is only relevant to product standards and reference standards.
- 对农产品监督抽查检验结果有异议的, 可以自收到检验报告之日起五日内, 向组织实施农产品质量安全监督抽查的农业行政主管部门或者其上级农业行政主管部门申请复检。对食品监督检验报告有异议的, 可以自收到检验报告之日起七个工作日内向实施抽样检验的食品药品监督管理部门或者其上一级食品药品监督管理部门提出复检申请, 对其它检验报告有异议的, 应于报告发出之日起十五日内向本院提出。  
 Any objections to the testing results of supervision sampling of agricultural products should apply for retest within 5 days upon receiving the test report to the administrative department of agriculture who organizes and implements agricultural products' supervision sampling or its superior department. Any objections to the testing results of supervision sampling of food should apply for retest within 7 days upon receiving the test report to the administrative department of food and drug who organizes and implements supervision sampling for food or its superior department. Any objections to other inspection report issued by SMQ should be submitted to SMQ within 15 days after the issuance of the test report.
- 电子版证书/报告更改后将不被追回, 委托方有义务将更改后的报告/证书提供给使用原报告/证书的相关方。  
 SMQ is not responsible for recalling the electronic version of the original report/certificate when any revision is made to them. The applicant assumes the responsibility of providing the revised version to any interested party who uses them.

投诉电话 Complaint hotline: 0755-86928949



## 检验报告

报告编号: WT201025604

第 2 页, 共 6 页

### 样品信息

样品名称: 防护口罩 KN95 Protective Face Mask  
 商标: HYNAUT 海氏海诺  
 型号/规格/等级: C004  
 样品编号: 200401  
 生产日期: 2020-04-09  
 生产单位: 青岛海诺生物工程有限公司  
 生产单位地址: 青岛市莱西市姜山镇工业园 海诺大厦  
 样品数量: 50只  
 抽样地点: \_\_\_\_\_  
 抽样人员: \_\_\_\_\_  
 检前样品描述: 正常, Normal.

抽样基数: \_\_\_\_\_

### 客户信息

委托单位: 青岛海诺生物工程有限公司 QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD.  
 委托单位地址: 青岛市莱西市姜山镇工业园广东路1号 NO.1 GUANGDONG ROAD, JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA  
 委托单位电话: 17561677800  
 邮政编码: \_\_\_\_\_  
 受检单位: \_\_\_\_\_

### 检验信息

委托日期: 2020年04月20日  
 检验类别: 送样检验  
 检验日期: 2020年04月20日 至 2020年04月27日  
 检验环境温度: (18~25) °C (30~70) °RH  
 判定依据: GB 2626-2006  
 检测依据: GB 2626-2006

委托单号: 8249572  
 获样方式: 送样

### 检验结论

检验结果见附页。  
 Test result refer to next page.

主检: 谢丹菊 谢丹菊 审核: 陈开江 陈开江



## 检验报告

报告编号: WT201025604

第 3 页, 共 6 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
1. 外观检查 Appearance (GB 2626-2006)	5.2条规定 5.2Item Requirement	1#~2#符合 Conformity	符合 Conformity
2. 过滤效率(%) Filtration efficiency (GB 2626-2006)	用氯化钠颗粒物检测: NaCl Non-oil aerosols KN95≥95.0	(GB 2626-2006) 未预处理样品 Unpretreated sample: 1#: 96.98 2#: 96.53 3#: 96.54 4#: 96.60 5#: 96.87 6#: 96.72 7#: 96.36 8#: 96.77 9#: 96.52 10#: 96.50  预处理样品 Pretreated sample: 1#: 96.33 2#: 96.60 3#: 96.57 4#: 96.19 5#: 96.47  KN类 KN-Series 温度 Temperature: 22.3°C 相对湿度 Relative humidity: 34.5% 颗粒物 Aerosol chamber: NaCl 颗粒物浓度 Concentration of aerosol chamber: 15mg/m <sup>3</sup> (检测流量 Flow meter rate: 85L/min)	符合 Conformity

## 检验报告

报告编号: WT204025604

第 4 页, 共 6 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
3. 呼吸阻力(Pa) Resistance of inhalation and exhalation (GB 2626-2006)	总吸气阻力Total Inhalation Resistance ≤350  总呼气阻力Total exhalation Resistance ≤250	(GB 2626-2006) 吸气阻力Inhalation resistance: 未预处理样品Unpretreated sample: 1# 2# 76.1 70.4  预处理样品Pretreated sample: 1# 2# 111.6 128.1  呼气阻力Exhalation resistance: 未预处理样品Unpretreated sample: 1# 2# 81.4 76.2  预处理样品Pretreated sample: 1# 2# 82.9 87.8  (通气量: 85L/min) (Flow: 85L/min)	符合 Conformity
4. 死腔(%) Dead space (GB 2626-2006)	≤1	(GB 2626-2006)  随弃式面罩Disposable facepiece: 0.59  (温度Temperature: 23.8℃)	符合 Conformity
5. 头带Head barness (GB 2626-2006)	随弃式面罩Disposable facepiece: 10N, 持续10s 不应出现滑脱、断裂 Noslippage, breakage	(GB 2626-2006)  未预处理样品Unpretreated sample: 1#~2#: 符合Pass  预处理样品Pretreatedsample: 1#~2#: 符合Pass	符合 Conformity

## 检验报告

报告编号: WT204025604

第 5 页, 共 6 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
6. 可燃性 Flammability (GB 2626-2006) 持续时间(s) Afterflame time	≤5	(GB 2626-2006)  随弃式面罩Disposable facepiece: 未预处理样品Unpretreated sample: 1# 2# 0.0 0.0  预处理样品Pretreated sample: 1# 2# 0.0 0.0  头模鼻尖位移速度: 60mm/s Displacement speed of Head mold nose tip: 60mm/s  距离燃烧器顶端20mm处的火焰温度: 815℃ Flame temperature at 20mm from the top of the burner: 815℃	符合 Conformity

附注:

1. 此报告以中文为准, 英文仅作参考。The Chinese version of this test report is the standard one, the English version is only for reference.  
2. 温度湿度预处理, 顺序按下述条件处理:

The sequence of temperature and humidity pretreatment is as follows:

- 在(38±2.5)℃和(85±5)%相对湿度环境放置(24±1)h  
Place at (38 ± 2.5) °C and (85 ± 5)% relative humidity for (24 ± 1) hours;
- 在(70±3)℃干燥环境处理(24±1)h  
Treatment in dry environment at (70 ± 3) °C for (24 ± 1) hours;
- (-30±3)℃环境放置(24±1)h  
Place at (-30 ± 3) °C for (24 ± 1) hours;

样品温度经恢复至室温后5h, 再进行检测  
After the sample temperature is restored to room temperature for 5 hours, the test shall be carried out again.

## 检验报告

报告编号: WT204025604

第 6 页, 共 6 页

3. 样品图片Photo(s) of the sample(s):



以下空白 END OF REPORT



中国认可  
国际互认  
检测  
TESTING  
CNAS L0599

Test Report **SL52025256537201TX** Date: May 21, 2020 Page 1 of 3  
QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD  
JIANGSHAN INDUSTRIAL PARK, LAIXI, QINGDAO, CHINA

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) Disposable Medical Face Mask  
Style No. : C010  
Composition : (A) PP Non-woven, Melt-blown fabric  
Sample Color : (A) Blue  
Manufacturer : QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD  
Country of Destination : EUR  
Supplier : QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD

Proposed Care Instruction : -

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : May 06, 2020  
Testing Period : May 09, 2020 - May 21, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Signed for and on behalf of  
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

*Sara Guo*

Sara Guo (Account Executive)



Test Report **SL52025256537201TX** Date: May 21, 2020 Page 2 of 3

Test Result

**Medical Face Masks-Requirements and Test Methods**  
(EN 14683:2019)

**Clause 5.2.2 Bacterial filtration efficiency (BFE)**  
(EN 14683 :2019 Annex B)

	1#	2#	3#	4#	5#
(BFE), %	>99.9	99.9	>99.9	99.9	99.9

Remark: Performance Requirement: Type I  $\geq 95\%$ , Type II  $\geq 98\%$ , Type IIR  $\geq 98\%$   
\* This test standard is not within the accredited scope in SGS Shanghai testing centre, it is carried out by external laboratory accredited by CMA (China Metrology Accreditation).

**Clause 5.2.3 Breathability (Differential Pressure)**  
(EN 14683 :2019 Annex C, Flow rate 8 l/min)

	1#	2#	3#	4#	5#
Differential pressure $\Delta P$ (Pa/cm <sup>2</sup> )	34	34	35	36	35

Remark: Performance Requirement: Type I <40 Pa/cm<sup>2</sup>, Type II <40 Pa/cm<sup>2</sup>, Type IIR <60 Pa/cm<sup>2</sup>

**Clause 5.2.4 Splash Resistance**  
(ISO 22609 :2004, Pressure 16.0 kPa)

Penetration on inside surface							
1#	2#	3#	4#	5#	6#	7#	8#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
9#	10#	11#	12#	13#	14#	15#	16#
Pass	Pass	Pass	Fail	Pass	Pass	Pass	Pass
17#	18#	19#	20#	21#	22#	23#	24#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
25#	26#	27#	28#	29#	30#	31#	32#
Pass	Pass	Pass	Pass	Pass	Pass	Pass	Pass
Number of Pass:	31						
Overall result:	Acceptable						

Remark:  
1) Performance Requirement Type I: N/A, Type II: N/A, Type IIR:  $\geq 16.0kPa$   
2) Distance of the medical face mask target area surface to the tip of cannula is  $300 \pm 10mm$ .  
3) Condition and Test temperature  $(21 \pm 5)^\circ C$ , relative humidity  $(85 \pm 10)\%$   
4) An acceptable quality limit of 4.0% is met for a single sampling plan when 29 or more of the 32 tested specimens show pass results




Test Report **SL52025256537201TX** Date: May 21, 2020 Page 3 of 3

**Clause 5.2.5 Microbial Cleanliness**  
(EN 14683: 2019 Annex D)

CFU/g	1#	2#	3#	4#	5#
	<1	<1	<1	<1	<1

Remark: Performance Requirement: Type Is30 CFU/g, Type IIs30 CFU/g, Type IIR $\leq 30$  CFU/g

Sample Photo



The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*

## 检验报告

TEST REPORT



(替代WT204025656报告)

报告编号: WT204032177

第 1 页, 共 5 页

**委托单位:** 青岛海诺生物工程有限公司 QINGDAO HAINO BIOLOGICAL ENGINEERING CO., LTD.  
**委托单位地址:** 青岛市莱西市姜山镇工业园广东路1号 NO.1 GUANGDONG ROAD, JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA  
**样品名称:** 见附页 Refer To Next Pages  
**型号/规格/等级:** 非无菌 耳挂式平面型 17.5cm\*9.5cm  
**检验类别:** 送样检验  
**检验地点:** 龙华实验基地 Longhua Experimental Base

深圳市计量质量检测研究院  
(检验检测专用章)

批准人: 何行月

签发日期: 2020年04月20日

签名: 何行月

## 检验报告

报告编号: WT204032177

第 3 页, 共 5 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
1. 外观 Appearance (YY/T 0969-2013)	4.1 条 4.1 Requirement	(YY/T 0969-2013) 1#~3# 符合 Conformity	符合 Conformity
2. 结构与尺寸 Structure and size (YY/T 0969-2013)	1. 结构 Structure 4.2 条 4.2 Requirement	(YY/T 0969-2013) 1#~3# 符合 Conformity	符合 Conformity
2. 尺寸 Size 4.2 条 4.2 Requirement	偏差 Deviation rate (%) 1# 2# 3# 长度 length: -0.6 -0.6 -0.6 宽度 width: -1.1 -1.1 -1.1		
3. 鼻梁 Nose clip (YY/T 0969-2013)	4.3.1 条 4.3.1 Requirement 4.3.2 条 4.3.2 Requirement	(YY/T 0969-2013) 1#~3# 符合 Conformity 长度 length (cm): 1# 2# 3# 10.2 10.3 10.2	符合 Conformity
4. 口罩带 Mask string (YY/T 0969-2013)	4.4.1 条 4.4.1 Requirement 4.4.2 条 4.4.2 Requirement	(YY/T 0969-2013) 符合 Conformity 1#~3# 符合 Conformity (定负荷 Fixed load: 10N, 持续 continuous: 5s)	符合 Conformity

## 检验报告

报告编号: WT204032177

第 4 页, 共 5 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
5. 细菌过滤效率 (%) Bacterial filtration efficiency (BFE) (YY/T 0969-2013)	(YY 0469-2011) 1# 2# 3# 100 100 100	(YY 0469-2011) 1# 2# 3# 100 100 100	符合 Conformity
6. 通气阻力 (Pa/cm <sup>2</sup> ) Ventilation resistance (YY/T 0969-2013)	(YY/T 0969-2013) 1# 2# 3# 39.3 42.7 39.1 (气体流量 Gas flow: 8L/min)	(YY/T 0969-2013) 1# 2# 3# 39.3 42.7 39.1 (气体流量 Gas flow: 8L/min)	符合 Conformity
7. 微生物 Microorganisms (YY/T 0969-2013)	(GB 15979-2002)	(GB 15979-2002)	符合 Conformity
细菌菌落总数 (CFU/g) Total amount of bacterial colony	≤100	<4	
大肠菌群 Coliform group	不得检出 No detected	未检出 Not detected	
绿脓杆菌 Pseudomonas aeruginosa	不得检出 No detected	未检出 Not detected	
金黄色葡萄球菌 Staphylococcus aureus	不得检出 No detected	未检出 Not detected	
溶血性链球菌 Streptococcus hemolyticus	不得检出 No detected	未检出 Not detected	
真菌 Fungal colony	不得检出 No detected	未检出 Not detected	



LABORATORIO ANALISI  
Errata Corrigé n° 3434 of 08/05/2020 to this Report which deletes and replaces the previous  
Test Report 361223 Date 06/05/2020 Sample n° 697240

Messrs **QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD.**  
**NO.1 GUANGDONG ROAD,**  
**JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA**

Sample Identification: XIAN5001 - DISPOSABLE SURGICAL MASK DATE OF MANUFACTURE 2020/03/17 - EXPIRATION DATE 2023/03/17 - 200317 - MANUFACTURER HAINUO  
Sampling: Customer  
Sampling site: Customer Site  
Transport: External Representative - Room Temperature  
Storage: Room Temperature  
Recording Date: 20/04/2020  
Beginning Test Date: 20/04/2020  
End Test Date: 05/05/2020

## MEDICAL FACE MASKS. REQUIREMENTS AND TEST METHODS

Scope Evaluation of performance requirements of the medical face mask according to UNI EN 14683:2019

Document digitally signed in accordance with current legislation by Dott.ssa Sonia Giannone - Ordine Naz. Dei Biologi Albo Professionale N° 050063  
TECNAL s.r.l. - Via Castelfranco, 17/D - Loc. Bazzano 40053 VALSAMOGGIA (BO) - Tel 051 832915 - Fax 051 830802 www.tecnal.it - laboratorio@tecnal.it P.I. 00579881202 - C.F.02460570373 - R.E.A. n° 290231 - Ca. Soc. 26.000,00 i.v. - Registro Laboratori Autocontrollo Emilia Romagna N° 008/BO/005



LABORATORIO ANALISI  
Errata Corrigé n° 3434 of 08/05/2020 to this Report which deletes and replaces the previous  
Test Report 361223 Date 06/05/2020 Sample n° 697240

Performance requirements for medical face masks (UNI EN 14683:2019, § 5.2.7):

Table 1

Test	Type I <sup>a</sup>	Type II	Type IIR
Bacterial filtration efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential pressure (Pa/cm <sup>2</sup> )	< 40	< 40	< 60
Splash resistance pressure (kPa)	Not required	Not required	≥ 16,0
Microbial cleanliness (cfu/g)	≤ 30	≤ 30	≤ 30

<sup>a</sup> Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

Summary of the test results:

Table 2

Test	Result	Compliance by type		
		Type I	Type II	Type IIR
Bacterial filtration efficiency (BFE), (%)	94,6	Passed	Passed	Passed
Differential pressure (Pa/cm <sup>2</sup> )	33	Passed	Passed	Passed
Splash resistance pressure (kPa)	Not performed			
Microbial cleanliness (cfu/g)	4	Passed	Passed	Passed

Document digitally signed in accordance with current legislation by Dott.ssa Sonia Giannone - Ordine Naz. Dei Biologi Albo Professionale N° 050063  
TECNAL s.r.l. - Via Castelfranco, 17/D - Loc. Bazzano 40053 VALSAMOGGIA (BO) - Tel 051 832915 - Fax 051 830802 www.tecnal.it - laboratorio@tecnal.it P.I. 00579881202 - C.F.02460570373 - R.E.A. n° 290231 - Ca. Soc. 26.000,00 i.v. - Registro Laboratori Autocontrollo Emilia Romagna N° 008/BO/005



LABORATORIO ANALISI  
Errata Corrigé n° 3434 of 08/05/2020 to this Report which deletes and replaces the previous  
Test Report 361223 Date 06/05/2020 Sample n° 697240

### Bacterial filtration efficiency (BFE)

#### Principle and Normative References

Scope Evaluation of efficiency of the medical face mask material(s) as a barrier to bacterial penetration  
Normative References UNI EN 14683:2019

#### Experimental conditions

N° of medical face masks tested 5  
Dimensions of the test specimens 100 mm x 100 mm  
Size of the area tested 49 cm<sup>2</sup>  
Side of the test specimen facing the aerosol Inside  
Flow rate during testing 28,3 L/min  
Test strain *Staphylococcus aureus* ATCC 6538  
Specimen conditioning 21 ± 5 °C and 85 ± 5 % HR for 1h  
Incubation conditions 37 ± 2 °C for 20-52 h

#### Procedure:

Following procedures are performed for each of the five tested masks.  
A representative test specimen of at least 100 mm x 100 mm is obtained from each mask. The specimen is clamped between a six-stage cascade impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* ATCC 6538 is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency (B) of each test specimen is calculated, as percentage using the following formula:

$$B = (C - T) / C \times 100$$

Where:

C is the number of the total plate counts for the two positive control runs;  
T is the total plate count for the test specimen.

Document digitally signed in accordance with current legislation by Dott.ssa Sonia Giannone - Ordine Naz. Dei Biologi Albo Professionale N° 050063  
TECNAL s.r.l. - Via Castelfranco, 17/D - Loc. Bazzano 40053 VALSAMOGGIA (BO) - Tel 051 832915 - Fax 051 830802 www.tecnal.it - laboratorio@tecnal.it P.I. 00579881202 - C.F.02460570373 - R.E.A. n° 290231 - Ca. Soc. 26.000,00 i.v. - Registro Laboratori Autocontrollo Emilia Romagna N° 008/BO/005

## Results

Table 3 summarizes results obtained with the Bacterial filtration efficiency (BFE) test.

Table 3

Parameter Test Method	U.M.	Results
Positive controls	cfu	1700-3000
Negative control	cfu	0
<b>Bacterial filtration efficiency BFE, Sample 1</b> <i>UNI EN 14683:2019 (Annex B)</i>	%	99,5
<b>Bacterial filtration efficiency BFE, Sample 2</b> <i>UNI EN 14683:2019 (Annex B)</i>	%	99,7
<b>Bacterial filtration efficiency BFE, Sample 3</b> <i>UNI EN 14683:2019 (Annex B)</i>	%	99,5
<b>Bacterial filtration efficiency BFE, Sample 4</b> <i>UNI EN 14683:2019 (Annex B)</i>	%	99,6
<b>Bacterial filtration efficiency BFE, Sample 5</b> <i>UNI EN 14683:2019 (Annex B)</i>	%	99,8
<b>Bacterial filtration efficiency BFE, Mean</b> <i>UNI EN 14683:2019 (Annex B)</i>	%	99,6

cfu: colony forming units; Positive control: test run without test specimen; Negative control: test run without bacterial suspension.

## Breathability

### Principle and Normative References

Scope Evaluation of efficiency of the air permeability of the mask.  
Normative References UNI EN 14683:2019

### Experimental conditions

N° of medical face masks tested 5  
General location of the areas tested Front  
Size of the area tested 4,9 cm<sup>2</sup>  
Side of the test specimen facing the airflow Inside  
Air flow rate during testing 8 L/min  
Specimen conditioning 21 ± 5 °C and 85 ± 5 % HR for 4h

### Procedure:

Following procedures are performed for each of the five tested masks.

A device, which measures the differential pressure required to draw air through a specimen surface area of 4,9 cm<sup>2</sup> at a constant air flow rate of 8 L/min, is used to measure the air exchange pressure of the medical face mask material. A differential manometer is used to measure the differential pressure required to move air through the specimen surface area.

## Results

Table 4 summarizes results obtained with the differential pressure test.

Table 4

Parameter Test Method	U.M.	Results
<b>Differential pressure, Sample 1</b> <i>UNI EN 14683:2019 (Annex C)</i>	Pa/cm <sup>2</sup>	34,2
<b>Differential pressure, Sample 2</b> <i>UNI EN 14683:2019 (Annex C)</i>	Pa/cm <sup>2</sup>	31,3
<b>Differential pressure, Sample 3</b> <i>UNI EN 14683:2019 (Annex C)</i>	Pa/cm <sup>2</sup>	30,6
<b>Differential pressure, Sample 4</b> <i>UNI EN 14683:2019 (Annex C)</i>	Pa/cm <sup>2</sup>	32,8
<b>Differential pressure, Sample 5</b> <i>UNI EN 14683:2019 (Annex C)</i>	Pa/cm <sup>2</sup>	32,6
<b>Differential pressure, Mean</b> <i>UNI EN 14683:2019 (Annex C)</i>	Pa/cm <sup>2</sup>	33

## Microbial cleanliness (Bioburden)

### Principle and Normative References

Scope	Evaluation of microbial cleanliness performance requirements in medical face mask.
Normative References	UNI EN 14683:2019

### Experimental conditions

N° of medical face masks tested	5
Culture media:	
Extraction liquid (used volume)	1g/L Peptone, 5g/L NaCl, 2g/L Polysorbete 20 (300 mL)
Total viable aerobic microbial count	TSA (Tryptic Soy Agar)
Total yeasts and moulds count	SDCA (Sabouraud Dextrose Agar with Chloramphenicol)
Extraction method	Orbital shaker for 5 min at 250 rpm
Analytical method	Membrane filtration (pore size 0,45 µm)
Incubation conditions:	
Total viable aerobic microbial count	30 ± 1 °C for 3 days
Total yeasts and moulds count	25 ± 1 °C for 7 days
Correction factor determined by the bioburden recovery efficiency	1,24

### Procedure:

Following procedures are performed for each of the five tested masks. Weigh each mask prior testing. The full mask is aseptically removed from the packaging and placed in a sterile 500 ml bottle containing 300 ml of extraction liquid. The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm. After this extraction step, 100 ml of the extraction liquid is filtered through a 0,45 µm filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 ml aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDCA for yeasts and moulds enumeration. The plates are incubated for 3 days at 30 °C and 7 days at 25 °C for TSA and SDCA plates respectively. The total bioburden is expressed by addition of the TSA and SDCA counts adjusted by a correction factor calculated from the bioburden recovery efficiency.

Document digitally signed in accordance with current legislation by Dott.ssa Sonia Giannone - Ordine Naz. Dei Biologi Albo Professionale N° 050063

TECNAL s.r.l. - Via Castelfranco, 17/D - Loc. Bazzano 40053 VALSAMOGGIA (BO) - Tel 051 832915 - Fax 051 830802 www.tecnal.it - laboratorio@tecnal.it P.I. 00579881202 - C.F. 02460570373 - R.E.A. n° 290231 - Ca. Soc. 26.000,00 i.v. - Registro Laboratori Autocontrollo Emilia Romagna N° 008/BO/005

## Results

Table 5 summarizes results obtained with the microbial cleanliness (bioburden) tests.

Table 5

Parameter Test Method	U.M.	Results
Microbial cleanliness (Bioburden), Sample 1 ISO 11737-1:2018	cfu/mask <sup>#</sup>	11
Microbial cleanliness (Bioburden), Sample 2 ISO 11737-1:2018	cfu/mask <sup>#</sup>	< 7
Microbial cleanliness (Bioburden), Sample 3 ISO 11737-1:2018	cfu/mask <sup>#</sup>	< 7
Microbial cleanliness (Bioburden), Sample 4 ISO 11737-1:2018	cfu/mask <sup>#</sup>	22
Microbial cleanliness (Bioburden), Sample 5 ISO 11737-1:2018	cfu/mask <sup>#</sup>	11
Microbial cleanliness (Bioburden), Total ISO 11737-1:2018	cfu/g <sup>#</sup>	4

<sup>#</sup> Values adjusted by the bioburden correction factor.

## Test Notes

MP: Laboratory-developed method  
U.M.: Measurement Unit  
Parameter Note: parameter information

N.R.: Not Detectable

The "Limits" column shows the limits of quantification or detectability indicated with "LDQ or LDR", the legal limits and / or guide values agreed with the client (The value indicated, if expressed in round brackets ( ), is to be considered "Guide Value". Otherwise it is to be considered "Law Limit").  
Errata Corrigé/Supplemento: the change made to this test report are indicated in underlined and italic font.  
The result of the quantitative evidence on surfaces is obtained by recalculation performed on the basis of the measure declared by the person who performed the sampling.

Quantitative microbiological tests:

Quantitative microbiological tests are performed in single replication in accordance with ISO 7218: 2007 / Amd1:

2013. Expression of the result by matrix Food / Surfaces in accordance with ISO 7218: 2007 / Amd1: 2013:

Present <4 or Present <40 or Present <400: means "microorganisms present but less than 4 or 40 or 400 cfu / g or cfu / ml" which means that the number of colonies grown in the plate at the first useful dilution is between 1ufc and 3 ufc.

Estimates: the estimated expression means that the number of colonies grown in the plate at the first useful dilution is between 4 cfu and 9 cfu.

Expression of the result by Water matrix in compliance with ISO 8199:2018:

Present <3 or Present <300: means "Microorganisms present in the volume" taken into consideration but less than 3 or 300 cfu / Volume" which means that the number of colonies grown in the plate at the first useful dilution is between 1 or 2 cfu.

Estimates: the estimated expression means that the number of colonies grown in the plate at the first useful dilution is between 3 cfu and 9 cfu

Uncertainty of measurement (I.D.M.):

For microbiological parameters the extended uncertainty of measurement is expressed as confidence interval (lower limit - upper limit) with coverage factor k = 2 and with confidence level of 95%. Quantitative tests are performed in a single replica in accordance with ISO 7218: 2007 / Amd1: 2013.

For chemical parameters the extended uncertainty values refer to a 95% confidence interval and a coverage factor k = 2.

Documentation traceability:

The description of Laboratory-developed methods (MP), test procedures (PP) methods normed and Operating Procedures (P.O.) are at your disposal in the laboratory.

In laboratory are available all the documentation to trace the technicians who carried out the tests, as well as the sampling and transport.

The results included in this Test Report refer only to the sample tested. In case the sampling is not performed by our staff, the laboratory is not responsible for the sample information reported in this test report and the results refer only to the sample received. This Test Report may not be partially reproduced, unless Tecnal's written approval.

Document digitally signed in accordance with current legislation by Dott.ssa Sonia Giannone - Ordine Naz. Dei Biologi Albo Professionale N° 050063

TECNAL s.r.l. - Via Castelfranco, 17/D - Loc. Bazzano 40053 VALSAMOGGIA (BO) - Tel 051 832915 - Fax 051 830802 www.tecnal.it - laboratorio@tecnal.it P.I. 00579881202 - C.F. 02460570373 - R.E.A. n° 290231 - Ca. Soc. 26.000,00 i.v. - Registro Laboratori Autocontrollo Emilia Romagna N° 008/BO/005

Document digitally signed in accordance with current legislation by Dott.ssa Sonia Giannone - Ordine Naz. Dei Biologi Albo Professionale N° 050063

TECNAL s.r.l. - Via Castelfranco, 17/D - Loc. Bazzano 40053 VALSAMOGGIA (BO) - Tel 051 832915 - Fax 051 830802 www.tecnal.it - laboratorio@tecnal.it P.I. 00579881202 - C.F. 02460570373 - R.E.A. n° 290231 - Ca. Soc. 26.000,00 i.v. - Registro Laboratori Autocontrollo Emilia Romagna N° 008/BO/005

## 检验报告

TEST REPORT



报告编号: WT204025664

第 1 页, 共 4 页

委托单位: 青岛海诺生物工程有限公司 QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD  
 委托单位地址: 青岛市莱西市姜山镇工业园广东路1号NO.1 GUANGDONG ROAD, JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA  
 样品名称: 防护口罩Protective Face Mask  
 型号/规格/等级: 无纺布平面型C001, C006  
 检验类别: 送样检验  
 检验地点: 龙华实验基地Longhua Experimental Base

深圳市计量质量检测研究院 (检验检测专用章) 批准人: 何行月 冯晶

签发日期: 2020年05月08日 签名: \_\_\_\_\_

## 检验报告

报告编号: WT204025664

第 2 页, 共 4 页

### 样品信息:

样品名称: 防护口罩Protective Face Mask  
 商标: HAINUO海氏海诺  
 型号/规格/等级: 无纺布平面型C001, C006  
 样品编号/批号: 200331  
 生产日期: 2020-03-31  
 生产单位: 青岛海诺生物工程有限公司  
 生产单位地址: 青岛市莱西市姜山镇工业园  
 样品数量: 50只  
 抽样地点: \_\_\_\_\_  
 抽样人员: \_\_\_\_\_  
 检前样品描述: 正常, Normal.

抽样基数: \_\_\_\_\_

### 客户信息:

委托单位: 青岛海诺生物工程有限公司 QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD  
 委托单位地址: 青岛市莱西市姜山镇工业园广东路1号NO.1 GUANGDONG ROAD, JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA  
 委托单位电话: 17561677800  
 邮政编码: \_\_\_\_\_  
 受托单位: \_\_\_\_\_

### 检验信息:

委托日期: 2020年04月20日 委托单号: 8249571  
 检验类别: 送样检验 获样方式: 送样  
 检验日期: 2020年04月20日至2020年05月08日  
 检验环境条件: (18~25)°C (20~70)%RH  
 判定依据: T/CTCA 7-2019  
 检测依据: 见附录 refer to next pages

### 检验结论:

检验结果见附录  
 Test result refer to next pages.

制: 胡月菊 审核: 陈开江 黄建飞  
 主检: \_\_\_\_\_ 审核: \_\_\_\_\_

## 检验报告

报告编号: WT204025664

第 3 页, 共 4 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
1. 通气阻力 (Pa) Ventilation resistance (T/CTCA 7-2019)	≤80	(YY 0469-2011) I# 2# 3# 34.4 33.7 37.8 空气流量Flow: 30L/min	符合 Conformity
2. 颗粒物过滤效率 Particle filtration efficiency (%) (T/CTCA 7-2019)	≥80	(YY 0469-2011) I# 2# 3# 94.23 93.98 94.66 气溶胶颗粒Aerosol chamber: NaCl 气溶胶浓度concentration of Aerosol chamber: 15mg/m <sup>3</sup> 温度Temperature: 24.7°C 相对湿度Relative humidity: 34.7%	符合 Conformity
3. 口罩带及口罩带与口罩体的连接处断裂强力 Tensile Strength of mask string and connect part between mask string and mask (N) (T/CTCA 7-2019)	≥10	(YY 0469-2011) I#-3#符合 Conformity (定负荷:10N, 持续5s) (fixed load:10N, Lasting for 5s)	符合 Conformity

## 检验报告

报告编号: WT204025664

第 4 页, 共 4 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
4. 细菌过滤效率 (BFE) (T/CTCA 7-2019)	≥95	(YY 0469-2011) I# 2# 3# 100 100 100	符合 Conformity

### 附注Note:

1. 此报告以中文为准, 英文仅作参考。The Chinese version of this test report is the standard one, the English version is only for reference.  
 2. 我院获CNAS认可能力范围未涉及T/CTCA 7-2019;  
 Testing scopes of CNAS do not involve T/CTCA 7-2019.



以下空白 END OF REPORT



## 检验报告

TEST REPORT



报告编号: WT204032179  
(替代WT204029174报告)

第 1 页, 共 4 页

**委托单位:** 青岛海诺生物工程有限公司 QINGDAO HAINUO BIOLOGICAL ENGINEERING CO., LTD.  
**委托单位地址:** 见附页 Refer To Next Pages  
**样品名称:** 防护口罩 KN95 Protective Face Mask  
**型号/规格/等级:** C005  
**检验类别:** 送样检验  
**检验地点:** 龙华实验基地 Longhua Experimental Base

深圳市计量质量检测研究院  
(检验检测专用章)

签发日期: 2020年04月30日

批准人: 何行月

签名: 何行月

## 检验报告

报告编号: WT204032179

第 3 页, 共 4 页

检验项目 Test Item	标准要求 Requirement	实测结果 Test Result	单项结论 Conclusion
死腔(% Dead space (GB 2626-2006)	≤1	随弃式面罩 Disposable facepiece: 1.56 (温度 Temperature: 23.8℃)	符合 Conformity

附注:

- 此报告以中文为准, 英文仅作参考, The Chinese version of this test report is the standard one, the English version is only for reference.
- 温湿度预处理, 顺序按下述条件处理:  
The sequence of temperature and humidity pretreatment is as follows:  
a) 在(38±2.5)℃和(85±5)%相对湿度环境放置(24±1)h  
Place at (38 ± 2.5) °C and (85 ± 5)% relative humidity for (24 ± 1) hours;  
b) 在(70±3)℃干燥环境处理(24±1)h  
Treatment in dry environment at (70 ± 3) °C for (24 ± 1) hours;  
c) (-30±3)℃环境放置(24±1)h  
Place at (-30 ± 3) °C for (24 ± 1) hours;  
样品温度经恢复至室温后5h, 再进行检测  
After the sample temperature is restored to room temperature for 5 hours, the test shall be carried out again.


3. 生产单位地址 Manufacturer Address, 委托单位地址 Applicant Address:  
青岛市莱西市姜山镇工业园广东路1号  
NO.1 GUANGDONG ROAD, JIANGSHAN INDUSTRIAL ZONE LAIXI CITY, QINGDAO, CHINA.

## 检验报告

报告编号: WT204032179

第 4 页, 共 4 页

4. 样品图片 Photo(s) of the sample(s):



以下空白 END OF REPORT



## 检验检测报告

### Test Report

(2020) WSZ FHL 第 1381 号

产品名称 防霾防护口罩  
 Product Name  
 委托单位 青岛海诺生物工程有限公司  
 Applicant  
 生产单位 青岛海诺生物工程有限公司  
 Manufacturer  
 检验检测类别 委托检验  
 Test Type

江苏国健检测技术有限公司  
 Jianguo Guojian Testing Technology Co., Ltd

## 检验检测报告

### Test Report

[2020] WSZ FHL 第 1381 号 共 2 页 第 1 页

产品名称 Product name	防霾防护口罩	规格型号 Specification	柳叶型 L
委托单位/地址/联系电话 Applicant/Addr/Tel	青岛海诺生物工程有限公司/青岛市莱西市姜山镇工业园/0532-86463333		
生产单位/地址/联系电话 Manufacturer/Addr/Tel	青岛海诺生物工程有限公司/青岛市莱西市姜山镇工业园/0532-86463333		
样品等级 Sample grade	KN95	样品编号 Sample number	GW 1381-2020
样品数量 Sample quantity	26 只	样品接收日期 Receiving date of sample	2020 年 02 月 28 日
检验检测类别 Test type	委托检验	货号/批号/款号 Article number/Batch number/Style number	批号: 200203
检验检测日期 Test date	2020/02/28~2020/03/02	检验检测地点 Test site	本公司检验室
样品状态 Sample state	符合检验检测要求		
检验检测依据 Test standard(s)	GB 2626-2006《呼吸防护用品 自吸过滤式防颗粒物呼吸器》		
检验检测项目 Test Items	过滤效率、呼吸阻力		
检验检测结论 Test conclusion	样品经检验, 所检项目符合 Q/0285HNS 039 标准规定的要求, 具体检验结果见第 2 页。 签发日期: 2020 年 3 月 5 日		
备注 Note	委托方要求所检项目按照 Q/0285HNS 039《防护口罩》进行判定 样品信息由委托方提供, 本报告仅对来样负责 报告有效期为 1 年		

批准: 陈 审核: 方恒 主检: 杨莹

## 检验检测结果

### Test Result

[2020] WSZ FHL 第 1381 号 共 2 页 第 2 页

序号 Number	检验检测项目 Test item	单位 Unit	技术要求 Technical requirement	检验检测结果 Test result	单项评价 Single item decision
1	呼吸阻力	Pa	≤175	未预处理	103.2
				预处理	98.7
	呼气阻力	Pa	≤145	未预处理	99.6
				预处理	97.3
2	过滤效率	—	≥95%	未预处理	83.1
				预处理	80.2
				未预处理	81.2
				预处理	80.7
				未预处理	98.9%
				预处理	98.8%
				未预处理	99.1%
				预处理	99.0%
				未预处理	99.1%
				预处理	99.2%
				未预处理	98.7%
				预处理	98.9%
未预处理	99.2%				
预处理	99.0%				
未预处理	98.7%				
预处理	98.8%				
未预处理	99.1%				
预处理	99.3%				

备注: 以下空白



# Partners



**HYNAUT**  
海氏海诺®



***Excellent products of Hainuo create a  
healthy and beautiful life.***